

---

## Contents

<b>Absorbance-Based Integrated Optical Sensors</b> M. Puyol · F. Villuendas · C. Domínguez · V. Cadarso A. Llobera · I. Salinas · I. Garcés · J. Alonso	1
<b>Luminescence Lifetime-Based Imaging of Sensor Arrays for High-Throughput Screening Applications</b> M. Schäferling	45
<b>Cataluminescence-Based Gas Sensors</b> M. Nakagawa · N. Yamashita	93
<b>Hollow Waveguide Infrared Spectroscopy and Sensing</b> C. M. Charlton · B. T. Thompson · B. Mizaikoff	133
<b>Combinatorial Method for Surface-Confining Sensor Design and Fabrication</b> L. Basabe-Desmonts · R. S. Zimmerman · D. N. Reinhoudt M. Crego-Calama	169
<b>The Interplay of Indicator, Support and Analyte in Optical Sensor Layers</b> G. Orellana · M. C. Moreno-Bondi · D. Garcia-Fresnadillo M. D. Marazuela	189
<b>Challenges in the Design of Optical DNA Biosensors</b> M. Massey · P. A. E. Piunno · U. J. Krull	227
<b>Gold Nanoparticles in Bioanalytical Assays and Sensors</b> N. T. K. Thanh · A. Vernhet · Z. Rosenzweig	261
<b>Reverse Symmetry Waveguide for Optical Biosensing</b> R. Horváth · N. Skivesen · N. B. Larsen · H. C. Pedersen	279

<b>Materials for Luminescent Pressure-Sensitive Paint</b>	
Y. Takeuchi · Y. Amao . . . . .	303
<b>Optical Sensing of Enantiomers</b>	
M. Kasper · S. Busche · G. Gauglitz . . . . .	323
<b>Optical Sensors for Ions and Protein Based on Digital Color Analysis</b>	
Y. Suzuki · K. Suzuki . . . . .	343
<b>Subject Index</b> . . . . .	367

## Absorbance-Based Integrated Optical Sensors

Mar Puyol<sup>1</sup> (✉) · Francisco Villuendas<sup>1</sup> · Carlos Domínguez<sup>2</sup> ·  
Víctor Cadarso<sup>2</sup> · Andreu Llobera<sup>3</sup> · Iñigo Salinas<sup>4</sup> · Ignacio Garcés<sup>4</sup> ·  
Julián Alonso<sup>5</sup>

<sup>1</sup>Dpto. Física Aplicada, Universidad de Zaragoza, Saragossa, Spain  
*mariadelmar.puyol@uab.es*

<sup>2</sup>Centro Nacional de Microelectrónica, Bellaterra, 08193 Barcelona, Spain

<sup>3</sup>Institut für Mikrotechnik, Technische Universität Braunschweig,  
Alte Salzdahlumer Str. 203, 38124 Braunschweig, Germany

<sup>4</sup>Dpto. Ingeniería Electrónica y Comunicaciones, Universidad de Zaragoza, CPS,  
María de Luna 3, 50015 Saragossa, Spain

<sup>5</sup>Grup de Sensors i Biosensors, Unitat de Química Analítica, Edifici Ciències, Bellaterra,  
08193 Barcelona, Spain

<b>1</b>	<b>Introduction</b>	<b>2</b>
<b>2</b>	<b>Design of Optochemical Sensors</b>	<b>4</b>
2.2	Reagents and Indicators	5
2.3	Supports and Matrices	6
2.3.1	Surface Optodes	7
2.3.2	Bulk Optodes	8
2.4	Sensor Configuration	9
2.4.1	Non-Guided Radiation Devices	9
2.4.2	Waveguide-Based Devices and Integrated Optochemical Sensors	12
2.5	Waveguide-Based Optochemical Sensors	18
2.5.1	Extrinsic Sensors	18
2.5.2	Intrinsic Sensors	19
<b>3</b>	<b>Absorbance-Based Integrated Devices</b>	<b>22</b>
3.1	Intrinsic Core-Based Optodes	22
3.2	Intrinsic Direct or Coating-Based Evanescent Wave Spectroscopic Optodes	25
3.3	Extrinsic Active Waveguides and IWAOs	26
3.4	IWAO Design and Construction	28
3.4.1	IWAOs Features	32
3.4.2	Technological Advantages	32
3.4.3	Sensitivity and Response Time	33
3.4.4	Applications	35
<b>References</b>		<b>40</b>

**Abstract** Optochemical sensors have had a huge expansion and in recent years interesting sorts of optical sensor have been developed which make use of the integrated circuit microelectronic technology and the optical technological advances achieved in the telecommunications industry. These devices are based on optical fibers, planar wave-

# Luminescence Lifetime-Based Imaging of Sensor Arrays for High-Throughput Screening Applications

Michael Schäferling

Institute of Analytical Chemistry, Chemo- and Biosensors, University of Regensburg,  
93040 Regensburg, Germany

*Michael.schaeferling@chemie.uni-regensburg.de*

<b>1</b>	<b>Introduction . . . . .</b>	<b>46</b>
<b>2</b>	<b>Lifetime-Imaging of Sensor Arrays Using Luminescent Probes . . . . .</b>	<b>47</b>
2.1	Basic Methods of Fluorescence Lifetime Imaging . . . . .	47
2.2	Imaging of Oxygen Partial Pressure . . . . .	52
2.3	Imaging of pH . . . . .	57
2.4	Imaging of Metal Ion and Anion Sensor Arrays . . . . .	61
2.5	Imaging of Hydrogen Peroxide . . . . .	64
2.6	Imaging of Citrate . . . . .	66
2.7	Imaging of Biomolecules Using the Europium(III) Tetracycline Probe: Determination of Glucose, Enzyme Activities, and ELISAs . . . . .	69
<b>3</b>	<b>Imaging of DNA and Protein Arrays with the Help of Luminescent Labels . . . . .</b>	<b>73</b>
3.1	Current Status of DNA and Protein Array Technology . . . . .	73
3.2	Imaging Techniques for Microarrays . . . . .	74
3.3	Luminescence Lifetime-Based Imaging of DNA and Protein Microarrays .	77
3.4	Resonance Energy Transfer Assays and Single Photon Counting Technology . . . . .	80
<b>4</b>	<b>Conclusion and Outlook . . . . .</b>	<b>84</b>
<b>References . . . . .</b>		<b>86</b>

**Abstract** This review highlights optical imaging technologies for the fluorescent read out of sensor arrays. Chemosensor arrays for the determination of pH, oxygen partial pressure or metal ions found particular applications in biomedical and environmental analysis. On the other hand, the monitoring of biomolecular interactions, e.g. of DNA sequences or proteins, is an important tool in pharmaceutical research and medical diagnosis. Microwell plate-based assays provided the possibility to analyze a large number of samples in parallel in a very short time. The development of microarray technologies was a step forward in miniaturization of high-throughput (or multiplexed) assay formats. The analysis of both microwell plate and microarray-based assays are subject of this survey, focussing on fluorescence lifetime imaging methods.

# Cataluminescence-Based Gas Sensors

Masuo Nakagawa<sup>1</sup> (✉) · Nobuhiko Yamashita<sup>2</sup>

<sup>1</sup>Faculty of Science, Okayama University of Science, 1-1 Ridai-cho, 700-0005 Okayama, Japan  
*masuo@dap.ous.ac.jp*

<sup>2</sup>Faculty of Education, Okayama University, 3-1-1 Tsushima-naka, 700-8530 Okayama, Japan  
*nobu@cc.okayama-u.ac.jp*

<b>1</b>	<b>Introduction</b>	94
<b>2</b>	<b>Background of CTL-Based Gas Sensors</b>	96
<b>3</b>	<b>Principles of CTL and Working Mechanism of CTL-Based Sensors</b>	97
3.1	Recombination Radiation	97
3.2	Radiation from the Excited Species	101
3.3	Working Mechanism of CTL-Based Sensors	104
3.3.1	Equation of CTL Intensity as a Function of Temperature in Laminar Flow	107
<b>4</b>	<b>Fabrication of Sensor Systems</b>	109
4.1	Sensor Materials	109
4.2	Sensor Cell	112
4.3	Sensor System	113
4.3.1	Fundamental Setup	113
<b>5</b>	<b>Characteristics of CTL-Based Gas Sensors</b>	115
5.1	CTL Spectrum	115
5.1.1	CTL Spectrum from the Excited Species	115
5.1.2	CTL Spectrum Due to Recombination Radiation	117
5.2	Temperature Dependence of CTL Intensity	118
5.3	Gas Flow-Rate Dependence of CTL Intensity	119
5.4	Gas Concentration Dependence of CTL Intensity	121
<b>6</b>	<b>Types of CTL-Based Sensing</b>	122
6.1	Continuous Measuring System	122
6.1.1	Discrimination of the Type of Vapor Based on the Difference in the CTL Spectrum	123
6.1.2	Recognition of Organic Vapors by Spectrum-Temperature Imaging	123
6.2	Accumulation-Type Measuring System	125
6.3	Temperature-Programming Type System	126
6.4	Multi-Sensor System	128
<b>7</b>	<b>Conclusions and Future Development</b>	130
<b>References</b>		132

# Hollow Waveguide Infrared Spectroscopy and Sensing

Christy M. Charlton · Bruce T. Thompson · Boris Mizaikoff (✉)

School of Chemistry and Biochemistry, Georgia Institute of Technology,  
Atlanta, GA 30332-0400, USA  
*boris.mizaikoff@chemistry.gatech.edu*

<b>1</b>	<b>Introduction</b>	134
1.1	Introduction to Infrared Gas Analysis	134
1.2	Qualitative IR Gas Analysis	135
1.3	Quantitative IR Gas Analysis	137
1.4	Fourier Transform IR Spectroscopy	138
<b>2</b>	<b>Hollow Waveguide Technology</b>	139
2.1	IR Waveguides	139
2.2	History of Hollow Waveguides	140
2.3	Hollow Waveguide Technology	142
2.3.1	Attenuated Total Reflection and Leaky Guides	142
2.3.2	Photonic Bandgap Fibers	142
2.4	Brief History of Hollow Waveguides as Gas-Transmission Cells	143
2.5	Optical Considerations for Hollow-Waveguide-Based Gas Sensing	144
<b>3</b>	<b>Applications of Fourier Transform IR Coupled Hollow Waveguide Gas Sensors</b>	147
3.1	Analyte Extraction from Intractable Sample Matrices	147
3.2	Real-Time Direct Fourier Transform IR Hollow Waveguide Gas Analysis on a Short Time Scale	154
<b>4</b>	<b>Applications of Laser-Coupled Hollow Waveguide Gas Sensors</b>	158
4.1	CO <sub>2</sub> Lasers Coupled to Hollow Waveguide Gas-Sensing Modules	159
4.2	Quantum Cascade Lasers Coupled to Hollow Waveguide Gas-Sensing Modules	160
4.3	Potential for Photonic Bandgap Fibers in Laser-Based Hollow Waveguide Gas Sensing	163
<b>5</b>	<b>Conclusion and Outlook</b>	163
<b>References</b>		165

**Abstract** Infrared (IR) hollow waveguides (HWGs) were first utilized in gas sensing applications in the early 1990's and have since been coupled to both FT-IR spectrometers and laser light sources. However, gas sensing with hollow waveguide modules has yet to achieve maturity for widespread use or device commercialization. We review this emerging field with emphasis on technology and application areas where we believe these devices are ideally suited, and discuss the advantages and limitations of using HWGs for chemical sensing, including their optical properties and waveguide losses. As new HWG technologies based on photonic bandgap materials are emerging, relevant applications of

# Combinatorial Method for Surface-Confining Sensor Design and Fabrication

Lourdes Basabe-Desmonts · Rebecca S. Zimmerman · David N. Reinhoudt · Mercedes Crego-Calama (✉)

Department of Supramolecular Chemistry and Technology, MESA<sup>+</sup> Institute for Nanotechnology and Faculty of Science and Technology, University of Twente, P.O. Box 217, 7500 AE Enschede, The Netherlands  
*m.cregocalama@utwente.nl*

1	Introduction	169
2	Fluorescent Monolayers on Glass as Probes Suitable for Combinatorial Methods	171
2.1	General Description	171
2.2	Fabrication Methods	173
2.2.1	Solution-Based Procedure	175
2.2.2	Microcontact Printing Based Procedure	183
2.3	Conclusions	186
	References	186

**Abstract** The procedure for the combinatorial fabrication of new sensing materials for cations and anions based on self-assembled monolayers (SAM) is discussed. A library of different sensitive substrates is generated by sequential deposition of fluorophores and small ligand molecules onto an amino-terminated SAM coated glass. The preorganization provided by the surface avoids the need for complex receptor design, allowing for a combinatorial approach to sensing systems based on individually deposited small molecules. Additionally the sensing system has been miniaturized to the microscale using microcontact printing and integrating the sensory SAMs on the walls of microchannels.

## 1 Introduction

Combinatorial methods are being widely implemented in the field of optical sensor development. The combinatorial concept is based on the relative ease of production of a large number of potential targets, of which it is hoped that some will exhibit the desired specifications of the researcher. This combinatorial approach is clearly different from the ‘classical’ rational design and individual creation of specific targets, in that the stress is not on the initial, specific design of the desired system, but on the testing of the large number of resultant targets to determine successful hits. Linked to a proper screening methodology and data processing, it allows for the facile search and

# The Interplay of Indicator, Support and Analyte in Optical Sensor Layers

Guillermo Orellana<sup>1</sup> (✉) · María C. Moreno-Bondi<sup>2</sup> ·  
David García-Fresnadillo<sup>1</sup> · María D. Marazuela<sup>2</sup>

<sup>1</sup>Department of Organic Chemistry, Faculty of Chemistry,  
Universidad Complutense de Madrid, 28040 Madrid, Spain  
*orellana@quim.ucm.es*

<sup>2</sup>Department of Analytical Chemistry, Faculty of Chemistry,  
Universidad Complutense de Madrid, 28040 Madrid, Spain  
*mcmbondi@quim.ucm.es*

1	Introduction	191
2	Organic Polymer Supports	192
2.1	Effect of the Support on the Indicator Response	193
2.1.1	Modulation of the Light Absorption/Emission Features of the Indicator	193
2.1.2	Linearity of the Response	195
2.2	Effect of the Polymer-Supported Indicator Concentration	195
2.3	Effect of the Support on the Analyte-Indicator Interaction	196
2.4	Effect of the Support on the Indicator Photostability	199
3	Inorganic Supports	200
3.1	Glass, Fused Silica and Silica Gels	200
3.2	Porous Glasses	201
3.3	Sol-Gel Materials	202
3.4	Zeolites	203
3.5	Effects of the Supported Indicator Heterogeneity	203
3.6	Some Examples	207
4	Biosensing Layers	209
4.1	Enzyme-Based Biosensors	211
4.1.1	Physical Entrapment	211
4.1.2	Covalent Binding	212
4.2	Whole-Cell Biosensors	213
4.3	Affinity Ligand-Based Biosensors	214
4.3.1	Immunosensors	214
4.3.2	Protein Microarrays	217
4.3.3	DNA Biosensors	219
5	Conclusions and Outlook	221
	References	222

**Abstract** It has been recognized since the pioneering times of fiber-optic sensing development that the best indicator dye is worth nothing without a (polymer) support fitted

## Challenges in the Design of Optical DNA Biosensors

Melissa Massey · Paul AE Piunno · Ulrich J Krull (✉)

Department of Chemical and Physical Sciences, University of Toronto at Mississauga,  
3359 Mississauga Rd. N., Mississauga, ON, L5L 1C6, Canada  
*ukrull@utm.utoronto.ca*

1	Introduction . . . . .	229
2	Considerations for the Preparation of Immobilized Nucleic Acid Films . . . . .	229
3	Development of the Selectivity of Optical Biosensors . . . . .	231
4	Optical Techniques for the Determination of Interfacial Nucleic Acid Hybridization . . . . .	233
4.1	Fibre Optic Sensors . . . . .	234
4.2	Surface Plasmon Resonance . . . . .	235
5	Particles as Labels for Transduction of Interfacial Target Binding . . . . .	237
6	Fluorescent Dyes for Detection of DNA Hybridization . . . . .	239
6.1	Fluorescent Labels . . . . .	239
6.2	Fluorescence from Dyes that Associate with DNA . . . . .	242
7	Fluorescence-Based Nucleic Acid Hybridization Assays . . . . .	245
7.1	Hybridization Assays . . . . .	245
7.2	Strategies for Fluorescence Transduction of Hybridization Based on Use of Labelled Binding Material . . . . .	248
7.2.1	Fluorescent Cationic Polymers . . . . .	248
7.2.2	Molecular Beacons . . . . .	248
7.2.3	Duplex Probes . . . . .	250
8	Detection of DNA Hybridization at Surfaces by Fluorescence . . . . .	251
9	Summary . . . . .	256
	References . . . . .	258

**Abstract** The field of biosensors and biochips for nucleic acid diagnostics has developed significantly over the last decade. High-throughput techniques offering the advantages of sensitivity and selectivity combined with rapid analysis to provide reproducible and accurate results are highly sought after in the areas of medical diagnostics, forensics, environmental monitoring, and bioterrorism. This chapter gives a short review of the necessary considerations for the preparation of immobilized nucleic acid films on a solid sensor substrate and the development of techniques utilized for the detection of selective hybridization of target binding materials. The fundamentals of fibre optic and surface plasmon resonance optical sensor platforms are outlined, followed by key developments

# Gold Nanoparticles in Bioanalytical Assays and Sensors

Nguyen Thi Kim Thanh<sup>1,2</sup> · Aude Vernhet<sup>3</sup> · Zeev Rosenzweig<sup>1</sup> (✉)

<sup>1</sup>Department of Chemistry, University of New Orleans, New Orleans, LA 70148, USA  
ntkthanj@liv.ac.uk, zrosenzw@uno.edu

<sup>2</sup>Centre for Nanoscale Science, Department of Chemistry & School of Biological Sciences, University of Liverpool, Crown Street, Liverpool L69 7ZD, UK  
ntkthanj@liv.ac.uk

<sup>3</sup>UMR Sciences pour l'Innovation, Campus ENSAM/INRA, bât. 28, Maître de Conférences ENSA-Montpellier, 2 place Viala, 34060 Montpellier, France

<b>1</b>	<b>Introduction</b>	262
<b>2</b>	<b>Experimental</b>	263
2.1	Materials and Reagents	263
2.2	Instrumentation	264
2.3	Synthesis of Gold Nanoparticles	264
<b>3</b>	<b>Results and Discussion</b>	265
3.1	Gold Nanoparticle Aggregation Immunoassays	265
3.2	Gold Nanoparticle Based pH Nanosensors	269
<b>4</b>	<b>Summary and Conclusions</b>	274
	<b>References</b>	276

**Abstract** In this review we report two major applications of gold nanoparticles in the field of bioassay and sensing. The first application is a unique, sensitive, and highly specific immunoassay system for antibodies using gold nanoparticles. The assay is based on the aggregation of gold nanoparticles that are coated with protein antigens in the presence of their corresponding antibodies. Aggregation of the gold nanoparticles results in an absorption change at 620 nm that is used to calibrate the amount of antibodies. The effects of pH, temperature, and the concentration of protein A-coated gold nanoparticles on the sensitivity of the assay were investigated. A dynamic range of two orders of magnitude and a limit of detection of 1 µg/mL of anti-protein A were observed.

The second application of nanoparticles is luminescence nanosensors, which have their potential use as site-specific probes in samples of limited dimensions. Novel methods of nanosensor fabrication to obtain nanosensors with improved analytical properties are reported. A new approach for controlled synthesis of fluorescence nanosensors for pH measurements is also presented. Gold nanoparticles were used as a supportive matrix for the sensing component. Polymer layers that include the active sensing element were deposited on the gold nanoparticles surface using an electrostatic-based layer by layer deposition method. Polymer layers of alternating charges were deposited on the particle surface through attractive electrostatic interactions. Such method enabled a more precise control of the size, size distribution and density of fluorophores on each particle. The study shows that this is an effective way to fabricate particle-based fluorescent nanosensors that are stable and effective in measuring the pH in aqueous media.

# Reverse Symmetry Waveguide for Optical Biosensing

Róbert Horváth (✉) · Nina Skivesen · Niels B. Larsen · Henrik C. Pedersen

Optics and Plasma Research Department, Risø National Laboratory, Building 128,  
Frederiksborgvej 399, P.O. Box 49, 4000 Roskilde, Denmark  
*robert.horvath@risoe.dk*

<b>1</b>	<b>Introduction</b>	279
<b>2</b>	<b>Waveguide Sensor Basics in Brief</b>	281
2.1	Basic Physics	281
2.2	Sensor Issues	284
2.3	Multimode Operation	287
<b>3</b>	<b>Reverse Symmetry Designs and Applications</b>	288
3.1	Nanoporous Waveguide	288
3.1.1	Refractometry	289
3.1.2	Bacterial Detection	291
3.1.3	Whole-Cell Detection	293
3.1.4	Amplitude Detection	294
3.2	Multimode Thin-Plate Waveguide	296
3.3	Microstructured Waveguide	297
<b>4</b>	<b>Summary</b>	298
<b>References</b>		300

**Abstract** The present chapter deals with a novel design of planar optical waveguide biosensors. The principle of reverse symmetry is based on making the refractive index (RI) of the waveguide substrate less than the RI of the medium covering the waveguiding film, which is usually an aqueous solution (RI  $\sim$  1.33). This is opposed to the conventional sensor geometry, where the substrate is glass or polymers with RIs of approximately 1.5. The reverse configuration can be used to tune the penetration depth of the evanescent electromagnetic field into the cover medium up to infinity; thus the waveguide can be tailor-made so that biological objects with any size can be probed by the evanescent field. This is an important improvement compared with, for example, surface plasmon resonance sensors, where the penetration depth is fixed by the choice of metal.

## 1 Introduction

For almost 20 years, the optical waveguide sensor has been used for the label-free, evanescent-field detection of chemical or biological reactions taking place in the close vicinity of the waveguide surface. This includes numerous applications within the fields of chemical and biological sensing, where in

## Materials for Luminescent Pressure-Sensitive Paint

Yumi Takeuchi · Yutaka Amao (✉)

Department of Applied Chemistry, Oita University, Dannoharu 700, 870-1192 Oita, Japan  
*amao@cc.oita-u.ac.jp*

1	Pressure-Sensitive Paint	303
2	Principles of PSP	304
3	Materials for Polymer-Based PSP	307
3.1	Coordination Compounds as Luminescent Probes for Polymer-Based PSP	311
3.2	Poly Aromatic Hydrocarbons as Luminescent Probes for Polymer-Based PSP	317
4	Materials for Probe Adsorbed Layer-Based PSP	318
	References	320

**Abstract** Pressure-sensitive paint (PSP) is applied to the aerodynamics measurement. PSP is optical sensor based on the luminescence of dye probe molecules quenching by oxygen gas. Many PSPs are composed of probe dye molecules, such as polycyclic aromatic hydrocarbons (pyrene, pyrene derivative etc.), transition metal complexes (ruthenium(II), osmium(II), iridium(III) etc.), and metalloporphyrins (platinum (II), palladium(II), etc.) immobilized in oxygen permeable polymer (silicone, polystyrene, fluorinated polymer, cellulose derivative, etc.) film. Dye probe molecules adsorbed layer based PSPs such as pyrene derivative and porphyrins directly adsorbed onto anodic oxidised aluminium plat substrate also developed. In this section the properties of various oxygen permeable polymer for matrix and various dye probes for PSP are described.

### 1

#### Pressure-Sensitive Paint

Surface pressure distribution measurement is of fundamental importance in the experimental study of aerodynamic problems in the fields of avionics, car, rocket, aerospace, and aircraft design [1]. The conventional methods based on pressure taps or transducers have a number of limitations. The most serious problem is that their very nature limits them to providing information only at discrete points on the surface of a substrate. A new approach to surface pressure distribution measurement, the use of pressure-sensitive paint (PSP), has recently developed that offers the potential of revolutionizing the nature of such measurements in the field of aerodynamics. This method employs the oxygen sensitivity of fluorescent materials in the form of a paint, in conjunction with image processing techniques, to map the pressure field over

## Optical Sensing of Enantiomers

Maura Kasper · Stefan Busche (✉) · Günter Gauglitz

Institute of Physical and Theoretical Chemistry, Eberhard Karls University,  
Auf der Morgenstelle 8, 72076 Tübingen, Germany  
*stefan.busche@ipc.uni-tuebingen.de*

<b>1</b>	<b>Introduction . . . . .</b>	<b>324</b>
<b>2</b>	<b>Principles of Discrimination of Enantiomers . . . . .</b>	<b>325</b>
2.1	Amide Phases . . . . .	325
2.2	Cyclodextrin Phases . . . . .	326
2.3	Molecularly Imprinted Polymers . . . . .	327
2.4	Fluorescence Sensors . . . . .	328
<b>3</b>	<b>Optical Sensing Applications for Enantiomers . . . . .</b>	<b>329</b>
3.1	Polymer-Based Sensors . . . . .	329
3.1.1	Chiral Amides . . . . .	329
3.1.2	Chiral Discrimination by Chirasil-Val . . . . .	329
3.1.3	Discrimination with Optical Chirasil-Calix Sensors . . . . .	332
3.1.4	Cyclodextrins . . . . .	333
3.1.5	Molecularly Imprinted Polymers . . . . .	335
3.2	Fluorescence Sensors . . . . .	336
3.2.1	Calixarene Fluorophores . . . . .	337
3.2.2	Fluorescent Cyclodextrins . . . . .	338
3.2.3	1,1'-Binaphthyl Fluorophores . . . . .	339
<b>4</b>	<b>Conclusions and Outlook . . . . .</b>	<b>340</b>
<b>References . . . . .</b>		<b>340</b>

**Abstract** The optical sensing of enantiomers is a current topic. During the last decade many different applications for optical sensing of enantiomers have been reported in the literature. The principles of distinction of enantiomers using amide and cyclodextrin phases, molecularly imprinted polymers and fluorescence sensors are depicted in this chapter. Label-free methods, like surface plasmon resonance and reflectometric interference spectroscopy, can be adopted for polymer-based chiral amides, cyclodextrines and molecularly imprinted polymers. These materials derived from chromatographic methods are used for enantiomeric separation in the aqueous phase as well as in the gaseous phase. Calixarene and 1,1'-binaphthyl fluorophores as well as fluorescent cyclodextrins could be established as sensing materials for fluorescence measurements. Examples for the most commonly used methods—enantioselective fluorescence quenching or enhancement—are presented.

# Optical Sensors for Ions and Protein Based on Digital Color Analysis

Yoshio Suzuki<sup>1,3</sup> · Koji Suzuki<sup>1,2</sup> (✉)

<sup>1</sup>Cooperation for Innovative Technology and Advanced Research in Evolutional Area (CITY AREA), Kanagawa Academy of Science and Technology, 3-2-1 Sakato, Takatsu-ku, Kawasaki, 213-0012 Kanagawa, Japan

<sup>2</sup>Department of Applied Chemistry, Faculty of Science and Technology, Keio University, 3-14-1 Hiyoshi, Kohoku-ku, Yokohama, 223-8522 Kanagawa, Japan  
*suzuki@aplc.keio.ac.jp*

<sup>3</sup>Present address:

National Institute of Advanced Industrial Science and Technology, 1-1-1 Higashi, Tsukuba, 305-8562 Ibaraki, Japan  
*suzuki-yoshio@aist.go.jp*

<b>1</b>	<b>Introduction</b>	<b>344</b>
<b>2</b>	<b><math>\text{Li}^+</math> Ion Sensors Based on Digital Color Analysis</b>	<b>345</b>
2.1	Theory	346
2.1.1	Response Mechanism for the Ion Optode	346
2.1.2	Theoretical Equations for Color Mixing on the Basis of $Q_x Q_y$ Coordinates	348
2.1.3	Optimization of the Mixing Ratio of the Dyes	350
2.2	Construction of a $\text{Li}^+$ Sensing Film Optode	352
2.3	Response of Optode Film 1 to $\text{Li}^+$ Concentration	353
2.4	Response of Optode Film 2 to $\text{Li}^+$ Concentration	355
<b>3</b>	<b><math>\text{NH}_4^+</math> Ion Sensor Based on Digital Color Analysis</b>	<b>357</b>
3.1	Construction of $\text{NH}_4^+$ Sensing Film Optode	358
3.2	Response of the $\text{NH}_4^+$ Film Optode	358
<b>4</b>	<b>Protein Sensor</b>	<b>361</b>
4.1	Simulation of the Color Change of the Mixed Dyes in Response to Human Serum Albumin	362
4.2	Actual Color Change of the Mixed Dyes in Response to Human Serum Albumin	363
<b>5</b>	<b>Conclusions and Future Direction</b>	<b>364</b>
<b>References</b>		<b>364</b>

**Abstract** A novel colorimetric method, digital color analysis (DCA), was proposed using a digital color analyzer and was applied to various quantitative analyses using chromaticity coordinates and suitable sensors for visual colorimetry based on the characteristics of human visual perception by virtual simulations based on digital color information. On the basis of DCA, we developed a visual colorimetric sensor for  $\text{Li}^+$ ,  $\text{NH}_4^+$  and protein determination by the mixing of two kinds of lipophilic dyes, whose optimum mixing ratio